

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060520\
 Data File : VU038594.D
 Acq On : 06 Jun 2020 02:09
 Operator : SY/MD
 Sample : L2910-01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9D02

Manual Integrations
 APPROVED

MMDadoda
 6/8/2020 1:47:43 PM

Quant Time: Jun 06 02:40:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 06 02:36:38 2020
 Response via : Initial Calibration



TIC: VU038594.D

(13) Acetone (T)

2.697min (+0.019) 0.91ug/L

response 3489

Ion	Exp%	Act%
43.05	100	100
58.05	32.80	18.86
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060520\
 Data File : VU038594.D
 Acq On : 06 Jun 2020 02:09
 Operator : SY/MD
 Sample : L2910-01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9D02

Manual Integrations
 APPROVED

MMDadoda
 6/8/2020 1:47:43 PM

Quant Time: Jun 06 02:40:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 06 02:36:38 2020
 Response via : Initial Calibration



(13) Acetone (T)

2.697min (+0.019) 2.59ug/L m

206/11/2020

response 9899

Ion	Exp%	Act%
43.05	100	100
58.05	32.80	6.65
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060520\
 Data File : VU038594.D
 Acq On : 06 Jun 2020 02:09
 Operator : SY/MD
 Sample : L2910-01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9D02

Manual Integrations
 APPROVED

MMDadoda
 6/8/2020 1:47:43 PM

Quant Time: Jun 06 03:28:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 06 02:36:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	296966	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	297812	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	148581	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	64675	3.37	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.40%
7) Chloroethane-d5	1.93	69	59796	4.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.00%
11) 1,1-Dichloroethene-d2	2.59	63	118405	3.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.60%
20) 2-Butanone-d5	4.69	46	318894	54.62	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.24%
24) Chloroform-d	5.10	84	163405	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
26) 1,2-Dichloroethane-d4	5.74	65	98557	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
32) Benzene-d6	5.76	84	326013	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.60%
36) 1,2-Dichloropropane-d6	6.72	67	102823	4.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.00%
41) Toluene-d8	7.92	98	275576	4.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.60%
43) trans-1,3-Dichloropropene-	8.20	79	39386	4.01	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.20%
46) 2-Hexanone-d5	8.66	63	269392	55.48	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.96%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	94635	4.88	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.60%
64) 1,2-Dichlorobenzene-d4	12.20	152	121178	4.73	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.60%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
12) 1,1-Dichloroethene	2.60	96	14699	0.804	ug/L	# 1
13) Acetone	2.70	43	9899m	2.591	ug/L	# 98
19) 1,1-Dichloroethane	3.91	63	83943	2.380	ug/L	# 98
27) 1,2-Dichloroethane	5.83	62	487594	19.329	ug/L	# 100
31) Carbon tetrachloride	5.56	117	3731	0.142	ug/L	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060520\
Data File : VU038594.D
Acq On : 06 Jun 2020 02:09
Operator : SY/MD
Sample : L2910-01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9D02

Manual Integrations
APPROVED

MMDadoda
6/8/2020 1:47:43 PM

Quant Time: Jun 06 03:28:24 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 06 02:36:38 2020
Response via : Initial Calibration

