

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060524\
 Data File : VU059153.D
 Acq On : 05 Jun 2024 17:14
 Operator : MD/SY
 Sample : VU0605WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0605WBL01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/10/2024
 Supervised By :Semsettin Yesilyurt 06/10/2024

Quant Time: Jun 06 03:46:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U060524DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Jun 06 03:28:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.113	96	39817	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.634	95	11667m	0.861	ug/l	0.00
Spiked Amount 1.000			Recovery	=	86.000%	
68) 1,2-Dichlorobenzene-d4	12.193	152	11292	0.796	ug/l	0.00
Spiked Amount 1.000			Recovery	=	80.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060524\
Data File : VU059153.D
Acq On : 05 Jun 2024 17:14
Operator : MD/SY
Sample : VU0605WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0605WBL01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/10/2024
Supervised By :Semsettin Yesilyurt 06/10/2024

Quant Time: Jun 06 03:46:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U060524DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Jun 06 03:28:28 2024
Response via : Initial Calibration

