

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060619\
 Data File : VU032479.D
 Acq On : 05 Jun 2019 22:24
 Operator : JC/SP
 Sample : K3213-17 50X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DB8K8

Quant Time: Jun 06 05:11:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 06 04:04:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	211825	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	199640	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	100833	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	51628	37.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.14%
7) Chloroethane-d5	1.67	69	45238	40.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.92%
11) 1,1-Dichloroethene-d2	2.27	63	80209	29.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.62%#
21) 2-Butanone-d5	4.17	46	108426	94.28	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.28%
24) Chloroform-d	4.64	84	113633	44.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.66%
26) 1,2-Dichloroethane-d4	5.30	65	76962	47.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.32%
32) Benzene-d6	5.33	84	240266	46.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.72%
36) 1,2-Dichloropropane-d6	6.32	67	76872	47.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.60%
41) Toluene-d8	7.56	98	228911	46.58	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.16%
43) trans-1,3-Dichloropropene-	7.85	79	34657	44.56	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.12%
47) 2-Hexanone-d5	8.31	63	72529	92.01	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.01%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	111083	47.17	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.34%
64) 1,2-Dichlorobenzene-d4	11.85	152	96254	49.84	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.68%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU060619\
Data File : VU032479.D
Acq On : 05 Jun 2019 22:24
Operator : JC/SP
Sample : K3213-17 50X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DB8K8

Quant Time: Jun 06 05:11:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 06 04:04:55 2019
Response via : Initial Calibration

