

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060619\
 Data File : VU032503.D
 Acq On : 06 Jun 2019 17:43
 Operator : JC/SP
 Sample : K3213-13
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DB8K3

Quant Time: Jun 07 06:24:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 07 05:48:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	215616	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	199397	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	101882	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	48614	34.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.50%
7) Chloroethane-d5	1.68	69	43345	38.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	77.12%
11) 1,1-Dichloroethene-d2	2.27	63	79423	29.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.00%#
21) 2-Butanone-d5	4.17	46	113983	97.37	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.37%
24) Chloroform-d	4.64	84	116156	44.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.04%
26) 1,2-Dichloroethane-d4	5.30	65	79067	48.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.22%
32) Benzene-d6	5.33	84	241648	46.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.36%
36) 1,2-Dichloropropane-d6	6.32	67	78028	48.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.14%
41) Toluene-d8	7.56	98	232710	47.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.82%
43) trans-1,3-Dichloropropene-	7.85	79	35568	45.79	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.58%
47) 2-Hexanone-d5	8.31	63	77062	97.88	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.88%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	114679	48.76	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.52%
64) 1,2-Dichlorobenzene-d4	11.85	152	97725	50.08	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.16%

Target Compounds					Ovalue
20) cis-1,2-Dichloroethene	4.22	96	1652	1.053 ug/L	92
34) Trichloroethene	6.19	95	7989	5.241 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060619\
Data File : VU032503.D
Acq On : 06 Jun 2019 17:43
Operator : JC/SP
Sample : K3213-13
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DB8K3

Quant Time: Jun 07 06:24:38 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 07 05:48:54 2019
Response via : Initial Calibration

