

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060619\
 Data File : VU032464.D
 Acq On : 05 Jun 2019 16:40
 Operator : JC/SP
 Sample : VU0606MBL01
 Misc : 5.00µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK46

Quant Time: Jun 06 04:14:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 06 04:04:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	203343	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	191182	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	96148	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	62051	47.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.06%
7) Chloroethane-d5	1.68	69	51383	48.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.94%
11) 1,1-Dichloroethene-d2	2.27	63	91782	35.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.06%
21) 2-Butanone-d5	4.17	46	111290	100.80	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.80%
24) Chloroform-d	4.64	84	115101	46.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.56%
26) 1,2-Dichloroethane-d4	5.30	65	75953	49.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.00%
32) Benzene-d6	5.33	84	240124	48.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.76%
36) 1,2-Dichloropropane-d6	6.32	67	75875	48.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.50%
41) Toluene-d8	7.56	98	226000	48.02	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.04%
43) trans-1,3-Dichloropropene-	7.85	79	36719	49.30	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.60%
47) 2-Hexanone-d5	8.31	63	72243	95.70	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.70%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	106547	47.25	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.50%
64) 1,2-Dichlorobenzene-d4	11.85	152	93652	50.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.70%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU060619\
Data File : VU032464.D
Acq On : 05 Jun 2019 16:40
Operator : JC/SP
Sample : VU0606MBL01
Misc : 5.00µg/5mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK46

Quant Time: Jun 06 04:14:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 06 04:04:55 2019
Response via : Initial Calibration

