

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060619\
 Data File : VU032472.D
 Acq On : 05 Jun 2019 19:43
 Operator : JC/SP
 Sample : K3213-10 50X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DB8K0

Quant Time: Jun 06 04:55:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 06 04:04:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	202955	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	190677	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	97874	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	50816	38.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.18%
7) Chloroethane-d5	1.67	69	45186	42.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.40%
11) 1,1-Dichloroethene-d2	2.27	63	81789	31.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.44%
21) 2-Butanone-d5	4.17	46	107397	97.46	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.46%
24) Chloroform-d	4.64	84	112401	45.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.54%
26) 1,2-Dichloroethane-d4	5.30	65	76299	49.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.64%
32) Benzene-d6	5.33	84	238005	48.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.16%
36) 1,2-Dichloropropane-d6	6.32	67	75718	48.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.56%
41) Toluene-d8	7.56	98	229445	48.88	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.76%
43) trans-1,3-Dichloropropene-	7.85	79	35308	47.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.06%
47) 2-Hexanone-d5	8.31	63	72322	96.06	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.06%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	111056	49.38	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.76%
64) 1,2-Dichlorobenzene-d4	11.86	152	97183	51.84	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.68%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU060619\
Data File : VU032472.D
Acq On : 05 Jun 2019 19:43
Operator : JC/SP
Sample : K3213-10 50X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DB8K0

Quant Time: Jun 06 04:55:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 06 04:04:55 2019
Response via : Initial Calibration

