

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060619\
 Data File : VU032484.D
 Acq On : 06 Jun 2019 09:25
 Operator : JC/SP
 Sample : VU0606MBL02
 Misc : 5.00µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK48

Quant Time: Jun 07 04:13:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 07 04:06:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	205592	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	191032	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	96730	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	60935	45.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.36%
7) Chloroethane-d5	1.68	69	51054	47.63	ug/L	0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.26%
11) 1,1-Dichloroethene-d2	2.27	63	92006	35.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.46%
21) 2-Butanone-d5	4.17	46	108395	97.11	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.11%
24) Chloroform-d	4.64	84	116574	46.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.72%
26) 1,2-Dichloroethane-d4	5.30	65	77297	49.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.64%
32) Benzene-d6	5.33	84	245505	49.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.00%
36) 1,2-Dichloropropane-d6	6.32	67	75800	48.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.48%
41) Toluene-d8	7.56	98	228376	48.57	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.14%
43) trans-1,3-Dichloropropene-	7.85	79	35849	48.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.34%
47) 2-Hexanone-d5	8.31	63	72075	95.55	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.55%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	109026	48.39	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.78%
64) 1,2-Dichlorobenzene-d4	11.85	152	94011	50.74	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.48%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU060619\
Data File : VU032484.D
Acq On : 06 Jun 2019 09:25
Operator : JC/SP
Sample : VU0606MBL02
Misc : 5.00µ/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK48

Quant Time: Jun 07 04:13:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 07 04:06:44 2019
Response via : Initial Calibration

