

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060622\
 Data File : VU048877.D
 Acq On : 06 Jun 2022 13:04
 Operator : SY/MD
 Sample : VU0606WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0606WBL01

Quant Time: Jun 07 05:14:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	271072	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	448455	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	408162	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	178059	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	201142	49.409	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.820%
35) Dibromofluoromethane	5.285	113	149858	46.840	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.680%
50) Toluene-d8	7.896	98	543850	47.421	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.840%
62) 4-Bromofluorobenzene	10.632	95	194485	44.840	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060622\
Data File : VU048877.D
Acq On : 06 Jun 2022 13:04
Operator : SY/MD
Sample : VU0606WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0606WBL01

Quant Time: Jun 07 05:14:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

