

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060622\
 Data File : VU048896.D
 Acq On : 06 Jun 2022 20:38
 Operator : SY/MD
 Sample : N3170-06RE
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 20220424-SPLP-AMENDED-COMPOSITE-DRE

Quant Time: Jun 07 05:21:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	256677	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	422662	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	388417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	176025	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	197167	51.148	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.300%
35) Dibromofluoromethane	5.288	113	66488	22.050	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	44.100%#
50) Toluene-d8	7.896	98	514331	47.584	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.160%
62) 4-Bromofluorobenzene	10.632	95	186478	45.617	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.240%
Target Compounds						
					Qvalue	
16) Acetone	2.623	43	70458	35.866	ug/l	100
20) Methylene Chloride	3.038	84	4674	1.399	ug/l	94
25) 2-Butanone	4.706	43	13670	4.844	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060622\
Data File : VU048896.D
Acq On : 06 Jun 2022 20:38
Operator : SY/MD
Sample : N3170-06RE
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
20220424-SPLP-AMENDED-COMPOSITE-DRE

Quant Time: Jun 07 05:21:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

