

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060622\
 Data File : VU048897.D
 Acq On : 06 Jun 2022 21:01
 Operator : SY/MD
 Sample : N3120-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 40032

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/07/2022
 Supervised By : Mahesh Dadoda 06/07/2022

Quant Time: Jun 07 05:22:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	253796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	418847	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	376993	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	180540	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	193438	50.751	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.500%			
35) Dibromofluoromethane	5.289	113	142842	47.803	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.600%			
50) Toluene-d8	7.896	98	501673	46.835	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 93.680%			
62) 4-Bromofluorobenzene	10.632	95	193960	47.880	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 95.760%			
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.183	59	19011	19.673	ug/l #	41
16) Acetone	2.620	43	3046626	1568.441	ug/l	99
18) Methyl Acetate	2.948	43	13068m	2.680	ug/l	
25) 2-Butanone	4.700	43	130004	46.588	ug/l	99
49) 1,4-Dioxane	6.970	88	3556	44.282	ug/l	88
59) 2-Hexanone	8.687	43	8876	2.081	ug/l	88
68) m/p-Xylenes	9.694	106	5457	1.038	ug/l	95
95) Naphthalene	14.089	128	9299	3.154	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060622\
Data File : VU048897.D
Acq On : 06 Jun 2022 21:01
Operator : SY/MD
Sample : N3120-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
40032

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/07/2022
Supervised By : Mahesh Dadoda 06/07/2022

Quant Time: Jun 07 05:22:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

