

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060719\
 Data File : VU032531.D
 Acq On : 07 Jun 2019 10:23
 Operator : JC/SP
 Sample : VU0607WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK52

Quant Time: Jun 07 22:25:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 07 22:19:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	216311	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	201391	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	102874	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	59540	42.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.84%
7) Chloroethane-d5	1.68	69	51022	45.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.48%
11) 1,1-Dichloroethene-d2	2.27	63	90093	32.79	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.58%
21) 2-Butanone-d5	4.17	46	117424	99.98	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.98%
24) Chloroform-d	4.64	84	120632	46.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.18%
26) 1,2-Dichloroethane-d4	5.30	65	79953	48.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.98%
32) Benzene-d6	5.33	84	250138	47.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.68%
36) 1,2-Dichloropropane-d6	6.32	67	79245	48.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.68%
41) Toluene-d8	7.56	98	232443	46.89	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.78%
43) trans-1,3-Dichloropropene-	7.85	79	36744	46.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.66%
47) 2-Hexanone-d5	8.31	63	77321	97.23	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.23%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	113230	47.67	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.34%
64) 1,2-Dichlorobenzene-d4	11.85	152	97304	49.38	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.76%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU060719\
Data File : VU032531.D
Acq On : 07 Jun 2019 10:23
Operator : JC/SP
Sample : VU0607WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK52

Quant Time: Jun 07 22:25:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 07 22:19:39 2019
Response via : Initial Calibration

