

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
 Data File : VU044063.D
 Acq On : 07 Jun 2021 18:19
 Operator : SY/MD
 Sample : VU0607WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0607WBL01

Quant Time: Jun 08 05:44:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 05:21:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	209778	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	419057	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	419474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	197304	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	192769	52.764	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.520%	
35) Dibromofluoromethane	5.298	113	149386	49.954	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.900%	
50) Toluene-d8	7.906	98	560335	48.577	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.160%	
62) 4-Bromofluorobenzene	10.639	95	205826	45.069	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.140%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
Data File : VU044063.D
Acq On : 07 Jun 2021 18:19
Operator : SY/MD
Sample : VU0607WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0607WBL01

Quant Time: Jun 08 05:44:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 05:21:34 2021
Response via : Initial Calibration

