

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
 Data File : VU044067.D
 Acq On : 07 Jun 2021 19:52
 Operator : SY/MD
 Sample : M2580-33
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B6(0-6)

Quant Time: Jun 08 05:45:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 05:21:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	213270	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	421986	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	443048	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	207853	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	196887	53.009	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.020%	
35) Dibromofluoromethane	5.298	113	147585	49.009	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.020%	
50) Toluene-d8	7.906	98	587060	50.541	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.080%	
62) 4-Bromofluorobenzene	10.639	95	223180	48.530	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.060%	
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	104177	48.597	ug/l	98
20) Methylene Chloride	3.054	84	10012	1.465	ug/l	93
25) 2-Butanone	4.716	43	33803	11.052	ug/l	98
43) Isopropyl Acetate	5.919	43	25852	3.053	ug/l	99
84) 1,2,4-Trimethylbenzene	11.475	105	4302	1.263	ug/l	68
95) Naphthalene	14.092	128	20715	4.224	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
Data File : VU044067.D
Acq On : 07 Jun 2021 19:52
Operator : SY/MD
Sample : M2580-33
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B6(0-6)

Quant Time: Jun 08 05:45:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 05:21:34 2021
Response via : Initial Calibration

