

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
 Data File : VU044068.D
 Acq On : 07 Jun 2021 20:16
 Operator : SY/MD
 Sample : M2580-34
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B6(6-12)

Quant Time: Jun 08 05:45:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 05:21:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	212424	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	417818	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	436869	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	200481	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	194584	52.597	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.200%	
35) Dibromofluoromethane	5.298	113	147153	49.353	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.700%	
50) Toluene-d8	7.906	98	579110	50.354	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.700%	
62) 4-Bromofluorobenzene	10.639	95	217264	47.715	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.420%	
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	16694	7.818	ug/l	95
20) Methylene Chloride	3.054	84	10776	1.725	ug/l	97
43) Isopropyl Acetate	5.919	43	12189	1.454	ug/l	98
95) Naphthalene	14.095	128	12402	3.701	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
Data File : VU044068.D
Acq On : 07 Jun 2021 20:16
Operator : SY/MD
Sample : M2580-34
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B6(6-12)

Quant Time: Jun 08 05:45:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 05:21:34 2021
Response via : Initial Calibration

