

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
 Data File : VU044069.D
 Acq On : 07 Jun 2021 20:39
 Operator : SY/MD
 Sample : M2612-21
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B12(6-10)

Quant Time: Jun 08 05:45:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 05:21:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	210260	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	417891	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	438532	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	202746	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	193383	52.811	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.620%	
35) Dibromofluoromethane	5.298	113	145754	48.875	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.760%	
50) Toluene-d8	7.906	98	580817	50.494	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.980%	
62) 4-Bromofluorobenzene	10.639	95	218852	48.055	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.120%	
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	42389	20.057	ug/l	97
17) Carbon Disulfide	2.797	76	9563	1.110	ug/l	97
20) Methylene Chloride	3.054	84	11497	1.996	ug/l	97
43) Isopropyl Acetate	5.919	43	28692	3.421	ug/l	100
95) Naphthalene	14.095	128	7193	3.336	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
Data File : VU044069.D
Acq On : 07 Jun 2021 20:39
Operator : SY/MD
Sample : M2612-21
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B12(6-10)

Quant Time: Jun 08 05:45:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 05:21:34 2021
Response via : Initial Calibration

