

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
 Data File : VU044070.D
 Acq On : 07 Jun 2021 21:03
 Operator : SY/MD
 Sample : M2612-22
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B12(10-14)

Quant Time: Jun 08 05:46:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 05:21:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	207948	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	417721	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	435176	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	202754	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	193897	53.540	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.080%	
35) Dibromofluoromethane	5.298	113	144920	48.616	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.240%	
50) Toluene-d8	7.906	98	575303	50.035	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.060%	
62) 4-Bromofluorobenzene	10.639	95	216623	47.585	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.180%	
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	63382	30.323	ug/l	99
20) Methylene Chloride	3.054	84	21162	5.230	ug/l	95
25) 2-Butanone	4.716	43	25071	8.407	ug/l	97
43) Isopropyl Acetate	5.919	43	25725	3.069	ug/l	99
95) Naphthalene	14.092	128	7823	3.379	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
Data File : VU044070.D
Acq On : 07 Jun 2021 21:03
Operator : SY/MD
Sample : M2612-22
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B12(10-14)

Quant Time: Jun 08 05:46:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 05:21:34 2021
Response via : Initial Calibration

