

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
 Data File : VU044073.D
 Acq On : 07 Jun 2021 22:13
 Operator : SY/MD
 Sample : M2612-25
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B12(22-26)

Quant Time: Jun 08 05:46:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 05:21:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	202284	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	418233	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	433167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	201988	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	192541	54.654	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.300%	
35) Dibromofluoromethane	5.298	113	142538	47.758	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.520%	
50) Toluene-d8	7.906	98	572592	49.738	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.480%	
62) 4-Bromofluorobenzene	10.639	95	217991	47.827	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.660%	
Target Compounds						
						Qvalue
16) Acetone	2.633	43	44023	21.651	ug/l	99
20) Methylene Chloride	3.051	84	10236	1.716	ug/l	94
43) Isopropyl Acetate	5.919	43	30093	3.585	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
Data File : VU044073.D
Acq On : 07 Jun 2021 22:13
Operator : SY/MD
Sample : M2612-25
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B12(22-26)

Quant Time: Jun 08 05:46:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 05:21:34 2021
Response via : Initial Calibration

