

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
 Data File : VU044081.D
 Acq On : 08 Jun 2021 01:21
 Operator : SY/MD
 Sample : M2580-28
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B4(2-4)

Quant Time: Jun 08 05:48:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 05:21:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	210838	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	421980	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	447571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	205814	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	193567	52.716	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.440%	
35) Dibromofluoromethane	5.298	113	146857	48.768	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.540%	
50) Toluene-d8	7.906	98	586927	50.530	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.060%	
62) 4-Bromofluorobenzene	10.639	95	220653	47.981	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.960%	
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	62505	29.494	ug/l	97
20) Methylene Chloride	3.054	84	9674	1.392	ug/l #	88
25) 2-Butanone	4.716	43	16449	5.440	ug/l	99
43) Isopropyl Acetate	5.922	43	23863	2.818	ug/l	99
95) Naphthalene	14.095	128	7389	3.342	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
Data File : VU044081.D
Acq On : 08 Jun 2021 01:21
Operator : SY/MD
Sample : M2580-28
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B4(2-4)

Quant Time: Jun 08 05:48:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 05:21:34 2021
Response via : Initial Calibration

