

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
 Data File : VU044084.D
 Acq On : 08 Jun 2021 02:31
 Operator : SY/MD
 Sample : M2580-31
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B5(2-4)

Quant Time: Jun 08 05:49:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 05:21:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	207576	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	419244	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	441474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	203595	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	193770	53.600	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.200%	
35) Dibromofluoromethane	5.298	113	147437	49.280	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.560%	
50) Toluene-d8	7.906	98	586898	50.858	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.720%	
62) 4-Bromofluorobenzene	10.639	95	212190	46.442	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.880%	
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	50534	24.220	ug/l	96
20) Methylene Chloride	3.054	84	10159	1.602	ug/l	98
43) Isopropyl Acetate	5.922	43	23651	2.811	ug/l	99
95) Naphthalene	14.095	128	24709	4.524	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060721\
Data File : VU044084.D
Acq On : 08 Jun 2021 02:31
Operator : SY/MD
Sample : M2580-31
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B5(2-4)

Quant Time: Jun 08 05:49:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 05:21:34 2021
Response via : Initial Calibration

