

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048905.D
 Acq On : 07 Jun 2022 13:27
 Operator : SY/MD
 Sample : N3165-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 59BR-20220531

Quant Time: Jun 08 04:51:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	265822	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	432569	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	397598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	178664	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	195340	48.931	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.860%		
35) Dibromofluoromethane	5.288	113	147896	47.924	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.840%		
50) Toluene-d8	7.899	98	524680	47.429	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	94.860%		
62) 4-Bromofluorobenzene	10.632	95	188178	44.979	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.960%		
Target Compounds						
16) Acetone	2.626	43	19014	9.346	ug/l	97
37) Ethyl Acetate	4.803	43	76372	13.413	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048905.D
Acq On : 07 Jun 2022 13:27
Operator : SY/MD
Sample : N3165-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
59BR-20220531

Quant Time: Jun 08 04:51:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

