

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048908.D
 Acq On : 07 Jun 2022 14:38
 Operator : SY/MD
 Sample : N3165-09DL 5X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 60BR-20220531DL

Quant Time: Jun 08 04:52:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	261049	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	425648	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	389529	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	177536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	195502	49.867	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.740%
35) Dibromofluoromethane	5.288	113	146963	48.396	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.800%
50) Toluene-d8	7.896	98	521280	47.888	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.780%
62) 4-Bromofluorobenzene	10.632	95	188094	45.690	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.380%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.665	96	35708	10.150	ug/l	89
37) Ethyl Acetate	4.803	43	166786	29.768	ug/l	99
44) Trichloroethene	6.542	130	179981	53.391	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048908.D
Acq On : 07 Jun 2022 14:38
Operator : SY/MD
Sample : N3165-09DL 5X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
60BR-20220531DL

Quant Time: Jun 08 04:52:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

