

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048913.D
 Acq On : 07 Jun 2022 16:36
 Operator : SY/MD
 Sample : N3165-15
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 09BR-20220601

Quant Time: Jun 08 04:53:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	265306	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	433747	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	388624	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	159515	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	197909	49.671	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.340%
35) Dibromofluoromethane	5.289	113	148402	47.957	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.920%
50) Toluene-d8	7.899	98	528095	47.608	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.220%
62) 4-Bromofluorobenzene	10.632	95	179126	42.699	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.400%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.604	62	2500	0.690	ug/l	94
16) Acetone	2.626	43	8228	4.052	ug/l	92
24) 1,1-Dichloroethane	3.867	63	10394	1.715	ug/l	95
27) cis-1,2-Dichloroethene	4.668	96	4683	1.310	ug/l	84
37) Ethyl Acetate	4.800	43	266946	46.755	ug/l	99
44) Trichloroethene	6.543	130	4917	1.431	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048913.D
Acq On : 07 Jun 2022 16:36
Operator : SY/MD
Sample : N3165-15
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
09BR-20220601

Quant Time: Jun 08 04:53:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

