

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048915.D
 Acq On : 07 Jun 2022 17:23
 Operator : SY/MD
 Sample : N3165-17
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 12S-20220601

Quant Time: Jun 08 04:54:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	247010	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	396186	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	356530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	158224	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	184425	49.715	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.440%
35) Dibromofluoromethane	5.288	113	139478	49.347	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.700%
50) Toluene-d8	7.899	98	487915	48.156	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.320%
62) 4-Bromofluorobenzene	10.632	95	171811	44.838	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.680%
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	7417	3.923	ug/l	97
37) Ethyl Acetate	4.800	43	129222	24.779	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048915.D
Acq On : 07 Jun 2022 17:23
Operator : SY/MD
Sample : N3165-17
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
12S-20220601

Quant Time: Jun 08 04:54:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

