

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048919.D
 Acq On : 07 Jun 2022 18:58
 Operator : SY/MD
 Sample : N3166-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB08-20220531

Quant Time: Jun 08 04:55:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	239890	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	409753	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	374047	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	166268	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	188478	52.316	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.640%
35) Dibromofluoromethane	5.289	113	136846	46.813	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.620%
50) Toluene-d8	7.899	98	498168	47.540	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.080%
62) 4-Bromofluorobenzene	10.632	95	178406	45.018	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.040%
Target Compounds						
16) Acetone	2.629	43	4023	2.191	ug/l	Qvalue # 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048919.D
Acq On : 07 Jun 2022 18:58
Operator : SY/MD
Sample : N3166-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB08-20220531

Quant Time: Jun 08 04:55:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

