

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048920.D
 Acq On : 07 Jun 2022 19:22
 Operator : SY/MD
 Sample : N3166-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 51BR-20220601

Quant Time: Jun 08 04:56:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	245781	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	400932	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	370979	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	161749	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	185014	50.124	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.240%
35) Dibromofluoromethane	5.288	113	140798	49.224	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.440%
50) Toluene-d8	7.899	98	480404	46.854	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.700%
62) 4-Bromofluorobenzene	10.632	95	170766	44.038	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.080%
Target Compounds						
					Qvalue	
16) Acetone	2.629	43	5692	3.026	ug/l	100
27) cis-1,2-Dichloroethene	4.668	96	891	0.269	ug/l	87
37) Ethyl Acetate	4.803	43	70625	13.382	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048920.D
Acq On : 07 Jun 2022 19:22
Operator : SY/MD
Sample : N3166-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
51BR-20220601

Quant Time: Jun 08 04:56:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

