

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048928.D
 Acq On : 07 Jun 2022 23:05
 Operator : SY/MD
 Sample : VU0607WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0607WBL02

Quant Time: Jun 08 04:58:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	245359	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	397510	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	370187	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	164975	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	186468	50.604	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.200%
35) Dibromofluoromethane	5.288	113	138101	48.697	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.400%
50) Toluene-d8	7.896	98	474435	46.670	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.340%
62) 4-Bromofluorobenzene	10.632	95	173054	45.012	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048928.D
Acq On : 07 Jun 2022 23:05
Operator : SY/MD
Sample : VU0607WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0607WBL02

Quant Time: Jun 08 04:58:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

