

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048930.D
 Acq On : 07 Jun 2022 23:52
 Operator : SY/MD
 Sample : N3164-15DL 5X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 38BR-20220531DL

Quant Time: Jun 08 09:13:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	253066	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	415895	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	382816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	163274	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	192360	50.614	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.220%
35) Dibromofluoromethane	5.288	113	144794	48.800	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.600%
50) Toluene-d8	7.899	98	508121	47.774	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.540%
62) 4-Bromofluorobenzene	10.632	95	179054	44.514	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.020%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.382	85	1338	0.462	ug/l	84
4) Vinyl Chloride	1.604	62	190417	55.127	ug/l	98
16) Acetone	2.629	43	3173	1.638	ug/l	98
27) cis-1,2-Dichloroethene	4.665	96	5701	1.672	ug/l	89
37) Ethyl Acetate	4.806	43	16396	2.995	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048930.D
Acq On : 07 Jun 2022 23:52
Operator : SY/MD
Sample : N3164-15DL 5X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
38BR-20220531DL

Quant Time: Jun 08 09:13:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

