

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048931.D
 Acq On : 08 Jun 2022 00:16
 Operator : SY/MD
 Sample : N3164-17DL 4X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 40BR-20220531DL

Quant Time: Jun 08 09:13:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	248164	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	415336	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	376999	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	168983	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	182678	49.015	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.040%
35) Dibromofluoromethane	5.288	113	142783	48.187	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.380%
50) Toluene-d8	7.896	98	484942	45.656	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.320%#
62) 4-Bromofluorobenzene	10.632	95	178959	44.550	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.100%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.385	85	1058	0.372	ug/l	80
4) Vinyl Chloride	1.604	62	27180	8.024	ug/l	94
16) Acetone	2.633	43	2783	1.465	ug/l #	85
21) trans-1,2-Dichloroethene	3.356	96	983	0.340	ug/l	92
27) cis-1,2-Dichloroethene	4.665	96	145865	43.617	ug/l	88
37) Ethyl Acetate	4.803	43	85360	15.613	ug/l	99
44) Trichloroethene	6.542	130	1368	0.416	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048931.D
Acq On : 08 Jun 2022 00:16
Operator : SY/MD
Sample : N3164-17DL 4X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
40BR-20220531DL

Quant Time: Jun 08 09:13:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

