

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048937.D
 Acq On : 08 Jun 2022 02:37
 Operator : SY/MD
 Sample : N3166-24
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 19BR-20220601

Quant Time: Jun 08 09:15:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	242945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	402701	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	358946	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	149371	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	185735	50.906	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.820%
35) Dibromofluoromethane	5.288	113	137112	47.725	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.460%
50) Toluene-d8	7.899	98	492517	47.824	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.640%
62) 4-Bromofluorobenzene	10.632	95	162095	41.618	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.240%
Target Compounds						Qvalue
16) Acetone	2.633	43	6489	3.490	ug/l	92
37) Ethyl Acetate	4.806	43	85754	16.178	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048937.D
Acq On : 08 Jun 2022 02:37
Operator : SY/MD
Sample : N3166-24
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
19BR-20220601

Quant Time: Jun 08 09:15:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

