

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048940.D
 Acq On : 08 Jun 2022 03:48
 Operator : SY/MD
 Sample : N3166-08 100X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP05-20220531

Quant Time: Jun 08 09:16:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	251730	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	400735	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	362664	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.813	152	168332	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	188935	49.976	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.960%
35) Dibromofluoromethane	5.289	113	141001	49.319	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.640%
50) Toluene-d8	7.896	98	495232	48.324	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.640%
62) 4-Bromofluorobenzene	10.633	95	169994	43.860	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.720%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.601	62	17428	5.072	ug/l	92
12) 1,1-Dichloroethene	2.578	96	635	0.248	ug/l	84
21) trans-1,2-Dichloroethene	3.353	96	2632	0.898	ug/l #	84
27) cis-1,2-Dichloroethene	4.665	96	238661	70.354	ug/l	88
44) Trichloroethene	6.543	130	9500	2.993	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048940.D
Acq On : 08 Jun 2022 03:48
Operator : SY/MD
Sample : N3166-08 100X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP05-20220531

Quant Time: Jun 08 09:16:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

