

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048941.D
 Acq On : 08 Jun 2022 04:11
 Operator : SY/MD
 Sample : N3166-26 100X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BRP-2-20220601

Quant Time: Jun 08 09:16:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	251724	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.247	114	401091	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	378554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	166108	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	191120	50.555	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.120%
35) Dibromofluoromethane	5.289	113	143010	49.978	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.960%
50) Toluene-d8	7.896	98	489293	47.702	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.400%
62) 4-Bromofluorobenzene	10.632	95	175155	45.152	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.300%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.382	85	1315	0.456	ug/l	99
4) Vinyl Chloride	1.604	62	50962	14.832	ug/l	100
21) trans-1,2-Dichloroethene	3.353	96	771	0.263	ug/l	85
27) cis-1,2-Dichloroethene	4.665	96	166060	48.953	ug/l	88
44) Trichloroethene	6.539	130	17743	5.586	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048941.D
Acq On : 08 Jun 2022 04:11
Operator : SY/MD
Sample : N3166-26 100X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BRP-2-20220601

Quant Time: Jun 08 09:16:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

