

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048943.D
 Acq On : 08 Jun 2022 04:58
 Operator : SY/MD
 Sample : N3166-20 100X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP08-20220601

Quant Time: Jun 08 09:17:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	241752	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	401062	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	374774	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	165151	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	185240	51.021	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.040%
35) Dibromofluoromethane	5.288	113	139770	48.849	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.700%
50) Toluene-d8	7.896	98	473700	46.185	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.360%
62) 4-Bromofluorobenzene	10.632	95	176901	45.605	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048943.D
Acq On : 08 Jun 2022 04:58
Operator : SY/MD
Sample : N3166-20 100X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP08-20220601

Quant Time: Jun 08 09:17:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

