

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048945.D
 Acq On : 08 Jun 2022 05:45
 Operator : SY/MD
 Sample : N3166-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP04-20220531

Quant Time: Jun 08 09:18:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	247952	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	404772	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	359241	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	154439	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	184137	49.449	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.900%
35) Dibromofluoromethane	5.289	113	136578	47.296	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.600%
50) Toluene-d8	7.899	98	474251	45.815	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.620%#
62) 4-Bromofluorobenzene	10.632	95	151321	38.653	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	77.300%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.385	85	5687	2.004	ug/l	99
4) Vinyl Chloride	1.604	62	112612	33.274	ug/l	99
12) 1,1-Dichloroethene	2.575	96	1017	0.403	ug/l	81
16) Acetone	2.629	43	6146	3.239	ug/l	92
21) trans-1,2-Dichloroethene	3.350	96	2790	0.966	ug/l	94
27) cis-1,2-Dichloroethene	4.668	96	594364	177.880	ug/l	88
37) Ethyl Acetate	4.803	43	330570	62.044	ug/l	99
44) Trichloroethene	6.542	130	5224	1.630	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048945.D
Acq On : 08 Jun 2022 05:45
Operator : SY/MD
Sample : N3166-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP04-20220531

Quant Time: Jun 08 09:18:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

