

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
 Data File : VU048947.D
 Acq On : 08 Jun 2022 06:32
 Operator : SY/MD
 Sample : N3164-20
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 42BR-20220531

Quant Time: Jun 08 09:18:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	249596	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	406838	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	369673	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	154685	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	190671	50.867	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.740%
35) Dibromofluoromethane	5.289	113	142514	49.101	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.200%
50) Toluene-d8	7.899	98	493820	47.463	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.920%
62) 4-Bromofluorobenzene	10.632	95	170638	43.366	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.740%
Target Compounds						
						Qvalue
16) Acetone	2.636	43	9218	4.825	ug/l	90
37) Ethyl Acetate	4.809	43	50911	9.507	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060722\
Data File : VU048947.D
Acq On : 08 Jun 2022 06:32
Operator : SY/MD
Sample : N3164-20
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
42BR-20220531

Quant Time: Jun 08 09:18:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

