

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060723\
 Data File : VU054423.D
 Acq On : 07 Jun 2023 20:09
 Operator : MD/SY
 Sample : 03075-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0FB1

Quant Time: Jun 08 01:24:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 08 01:20:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	291597	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	277013	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	136096	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	82024	4.420	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.400%
7) Chloroethane-d5	1.914	69	81466	4.689	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.800%
11) 1,1-Dichloroethene-d2	2.566	65	37049	4.470	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	89.400%
20) 2-Butanone-d5	4.631	46	271851	52.103	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.200%
24) Chloroform-d	5.058	84	184188	4.748	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.000%
26) 1,2-Dichloroethane-d4	5.698	65	94812	5.014	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.200%
32) Benzene-d6	5.724	84	363956	4.726	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
36) 1,2-Dichloropropane-d6	6.685	67	126787	4.930	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.600%
41) Toluene-d8	7.894	98	310165	4.539	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
43) trans-1,3-Dichloroprop...	8.177	79	40595	4.539	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.800%
46) 2-Hexanone-d5	8.630	63	178327	47.531	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.060%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	92864	4.842	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	115707	5.040	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060723\
Data File : VU054423.D
Acq On : 07 Jun 2023 20:09
Operator : MD/SY
Sample : 03075-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0FB1

Quant Time: Jun 08 01:24:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 08 01:20:50 2023
Response via : Initial Calibration

