

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060724\
 Data File : VU059186.D
 Acq On : 07 Jun 2024 14:55
 Operator : MD/SY
 Sample : P2696-19
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK002

Quant Time: Jun 08 03:17:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 08 01:56:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	208709	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	209950	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	90960	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	72846	58.294	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	116.580%
7) Chloroethane-d5	1.907	69	60814	56.811	ug/L	0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	113.620%
11) 1,1-Dichloroethene-d2	2.560	63	90344	36.627	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.260%
21) 2-Butanone-d5	4.611	46	105652	89.699	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.700%
24) Chloroform-d	5.055	84	121313	48.766	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.540%
26) 1,2-Dichloroethane-d4	5.695	65	79266	51.987	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.980%
32) Benzene-d6	5.721	84	249742	51.828	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.660%
36) 1,2-Dichloropropane-d6	6.682	67	83736	50.399	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.800%
41) Toluene-d8	7.894	98	214932	52.543	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.080%
43) trans-1,3-Dichloroprop...	8.177	79	33332	49.631	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.260%
47) 2-Hexanone-d5	8.627	63	58276	75.849	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.850%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	116793	41.535	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.080%
66) 1,2-Dichlorobenzene-d4	12.190	152	72443	53.368	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.740%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060724\
Data File : VU059186.D
Acq On : 07 Jun 2024 14:55
Operator : MD/SY
Sample : P2696-19
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK002

Quant Time: Jun 08 03:17:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 08 01:56:01 2024
Response via : Initial Calibration

