

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060724\
 Data File : VU059182.D
 Acq On : 07 Jun 2024 13:23
 Operator : MD/SY
 Sample : P2693-02 50X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHHT2

Quant Time: Jun 08 03:16:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 08 01:56:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	213297	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	215712	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	91753	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	41460	32.464	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	64.920%	
7) Chloroethane-d5	1.907	69	41527	37.959	ug/L	0.04
Spiked Amount	50.000	Range 70 - 130	Recovery	=	75.920%	
11) 1,1-Dichloroethene-d2	2.560	63	61983	24.588	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	49.180%#	
21) 2-Butanone-d5	4.611	46	118262	98.245	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	98.240%	
24) Chloroform-d	5.055	84	113214	44.531	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.060%	
26) 1,2-Dichloroethane-d4	5.695	65	71533	45.906	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.820%	
32) Benzene-d6	5.721	84	203672	41.138	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	82.280%	
36) 1,2-Dichloropropane-d6	6.685	67	75758	44.379	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	88.760%	
41) Toluene-d8	7.891	98	161734	38.482	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	76.960%#	
43) trans-1,3-Dichloroprop...	8.177	79	28757	41.675	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	83.360%	
47) 2-Hexanone-d5	8.627	63	63426	80.347	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	80.350%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	129050	44.668	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	89.340%	
66) 1,2-Dichlorobenzene-d4	12.190	152	65558	47.878	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	95.760%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060724\
Data File : VU059182.D
Acq On : 07 Jun 2024 13:23
Operator : MD/SY
Sample : P2693-02 50X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHHT2

Quant Time: Jun 08 03:16:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 08 01:56:01 2024
Response via : Initial Calibration

