

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060819\
 Data File : VU032543.D
 Acq On : 08 Jun 2019 15:20
 Operator : JC/SP
 Sample : K3276-17 50X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DB8F3

Quant Time: Jun 08 23:30:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 08 23:12:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	213316	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	195109	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	102825	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	74527	53.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	107.70%
7) Chloroethane-d5	1.68	69	58205	52.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	104.66%
11) 1,1-Dichloroethene-d2	2.27	63	107744	39.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.52%
21) 2-Butanone-d5	4.17	46	118101	101.97	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.97%
24) Chloroform-d	4.64	84	128041	49.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.22%
26) 1,2-Dichloroethane-d4	5.30	65	85321	52.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.94%
32) Benzene-d6	5.33	84	270504	53.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.80%
36) 1,2-Dichloropropane-d6	6.32	67	85145	53.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.22%
41) Toluene-d8	7.56	98	253064	52.69	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.38%
43) trans-1,3-Dichloropropene-	7.85	79	41701	54.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	109.72%
47) 2-Hexanone-d5	8.31	63	81397	105.66	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.66%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	114474	49.74	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.48%
64) 1,2-Dichlorobenzene-d4	11.85	152	102865	52.23	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.46%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU060819\
Data File : VU032543.D
Acq On : 08 Jun 2019 15:20
Operator : JC/SP
Sample : K3276-17 50X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DB8F3

Quant Time: Jun 08 23:30:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM060619WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 08 23:12:44 2019
Response via : Initial Calibration

