

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060820\
 Data File : VU038660.D
 Acq On : 08 Jun 2020 10:52
 Operator : SY/MD
 Sample : VU0608WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK07

Quant Time: Jun 09 00:52:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 09 00:46:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	323473	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	318361	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	147735	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	114628	5.48	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.60%
7) Chloroethane-d5	1.93	69	89543	5.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.40%
11) 1,1-Dichloroethene-d2	2.59	63	179312	4.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.00%
20) 2-Butanone-d5	4.69	46	385057	60.55	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.10%
24) Chloroform-d	5.10	84	214126	5.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.80%
26) 1,2-Dichloroethane-d4	5.74	65	130263	5.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	116.60%
32) Benzene-d6	5.76	84	446529	5.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.40%
36) 1,2-Dichloropropane-d6	6.72	67	134566	5.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.20%
41) Toluene-d8	7.92	98	376416	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.60%
43) trans-1,3-Dichloropropene-	8.20	79	56336	5.36	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	107.20%
46) 2-Hexanone-d5	8.66	63	317914	61.25	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	122.50%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	113674	5.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.60%
64) 1,2-Dichlorobenzene-d4	12.20	152	145322	5.70	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU060820\
Data File : VU038660.D
Acq On : 08 Jun 2020 10:52
Operator : SY/MD
Sample : VU0608WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK07

Quant Time: Jun 09 00:52:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 09 00:46:08 2020
Response via : Initial Calibration

