

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060820\
 Data File : VU038665.D
 Acq On : 08 Jun 2020 13:10
 Operator : SY/MD
 Sample : L2906-07
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Jun 09 01:23:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 09 00:46:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	322858	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	315337	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	149066	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	115506	5.53	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.60%
7) Chloroethane-d5	1.93	69	89055	5.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.00%
11) 1,1-Dichloroethene-d2	2.59	63	173031	4.21	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.20%
20) 2-Butanone-d5	4.68	46	375423	59.14	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.28%
24) Chloroform-d	5.10	84	212968	5.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.60%
26) 1,2-Dichloroethane-d4	5.74	65	124804	5.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.00%
32) Benzene-d6	5.76	84	443306	5.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.60%
36) 1,2-Dichloropropane-d6	6.72	67	129461	5.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.00%
41) Toluene-d8	7.92	98	379899	5.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.60%
43) trans-1,3-Dichloropropene-	8.20	79	55186	5.30	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	106.00%
46) 2-Hexanone-d5	8.65	63	317149	61.69	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	123.38%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	112563	5.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.60%
64) 1,2-Dichlorobenzene-d4	12.20	152	149831	5.83	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU060820\
Data File : VU038665.D
Acq On : 08 Jun 2020 13:10
Operator : SY/MD
Sample : L2906-07
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jun 09 01:23:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 09 00:46:08 2020
Response via : Initial Calibration

