

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
 Data File : VU044100.D
 Acq On : 08 Jun 2021 09:10
 Operator : SY/MD
 Sample : M2613-23
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B12(122-126)

Quant Time: Jun 08 23:10:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 17:38:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	207050	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	420671	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	439455	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	201454	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	192024	53.252	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.500%	
35) Dibromofluoromethane	5.298	113	145730	48.544	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.080%	
50) Toluene-d8	7.906	98	579136	50.015	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.020%	
62) 4-Bromofluorobenzene	10.639	95	215940	47.103	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.200%	
Target Compounds					Qvalue	
16) Acetone	2.632	43	13559	6.515	ug/l	98
20) Methylene Chloride	3.051	84	29347	7.975	ug/l	94
43) Isopropyl Acetate	5.919	43	27766	3.289	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
Data File : VU044100.D
Acq On : 08 Jun 2021 09:10
Operator : SY/MD
Sample : M2613-23
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B12(122-126)

Quant Time: Jun 08 23:10:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 17:38:51 2021
Response via : Initial Calibration

