

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
 Data File : VU044104.D
 Acq On : 08 Jun 2021 10:44
 Operator : SY/MD
 Sample : M2612-27
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B12(30-34)

Quant Time: Jun 08 23:11:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 17:38:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	211264	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	429757	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	451190	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	204711	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	195365	53.098	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.200%	
35) Dibromofluoromethane	5.298	113	149993	48.908	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.820%	
50) Toluene-d8	7.906	98	594420	50.249	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.500%	
62) 4-Bromofluorobenzene	10.639	95	217908	46.527	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.060%	
Target Compounds					Qvalue	
16) Acetone	2.633	43	19302	9.090	ug/l	97
20) Methylene Chloride	3.057	84	9496	1.328	ug/l	95
43) Isopropyl Acetate	5.919	43	26361	3.056	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
Data File : VU044104.D
Acq On : 08 Jun 2021 10:44
Operator : SY/MD
Sample : M2612-27
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B12(30-34)

Quant Time: Jun 08 23:11:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 17:38:51 2021
Response via : Initial Calibration

