

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
 Data File : VU044106.D
 Acq On : 08 Jun 2021 11:31
 Operator : SY/MD
 Sample : M2612-29
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B12(38-42)

Quant Time: Jun 08 23:12:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 17:38:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	211261	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	421232	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	454684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	211020	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	191601	52.076	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.160%	
35) Dibromofluoromethane	5.298	113	149678	49.793	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.580%	
50) Toluene-d8	7.906	98	597207	51.507	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.020%	
62) 4-Bromofluorobenzene	10.639	95	220588	48.052	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.100%	
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	18887	8.894	ug/l	97
20) Methylene Chloride	3.054	84	9061	1.187	ug/l	96
43) Isopropyl Acetate	5.919	43	26549	3.140	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
Data File : VU044106.D
Acq On : 08 Jun 2021 11:31
Operator : SY/MD
Sample : M2612-29
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B12(38-42)

Quant Time: Jun 08 23:12:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 17:38:51 2021
Response via : Initial Calibration

