

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
 Data File : VU044108.D
 Acq On : 08 Jun 2021 12:18
 Operator : SY/MD
 Sample : M2612-31
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B12(46-50)

Quant Time: Jun 08 23:12:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 17:38:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	206262	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	413897	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	444637	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	203593	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	188054	52.351	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.700%	
35) Dibromofluoromethane	5.298	113	145660	49.315	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.640%	
50) Toluene-d8	7.906	98	579940	50.904	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.800%	
62) 4-Bromofluorobenzene	10.639	95	218646	48.473	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.940%	
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	19515	9.413	ug/l	97
20) Methylene Chloride	3.051	84	9932	1.548	ug/l	94
43) Isopropyl Acetate	5.922	43	26769	3.223	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
Data File : VU044108.D
Acq On : 08 Jun 2021 12:18
Operator : SY/MD
Sample : M2612-31
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 58 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B12(46-50)

Quant Time: Jun 08 23:12:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 17:38:51 2021
Response via : Initial Calibration

