

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
 Data File : VU044111.D
 Acq On : 08 Jun 2021 13:28
 Operator : SY/MD
 Sample : M2612-34
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B12(58-62)

Quant Time: Jun 08 23:13:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 17:38:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	211522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.259	114	427857	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	454267	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	202527	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	195297	53.015	ug/l	0.00
Spiked Amount 50.000			Recovery	= 106.040%		
35) Dibromofluoromethane	5.298	113	151954	49.768	ug/l	0.00
Spiked Amount 50.000			Recovery	= 99.540%		
50) Toluene-d8	7.906	98	596777	50.673	ug/l	0.00
Spiked Amount 50.000			Recovery	= 101.340%		
62) 4-Bromofluorobenzene	10.639	95	215427	46.201	ug/l	0.00
Spiked Amount 50.000			Recovery	= 92.400%		
Target Compounds						
						Qvalue
16) Acetone	2.633	43	18255	8.586	ug/l	95
20) Methylene Chloride	3.057	84	9509	1.329	ug/l	94
43) Isopropyl Acetate	5.919	43	28056	3.267	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
Data File : VU044111.D
Acq On : 08 Jun 2021 13:28
Operator : SY/MD
Sample : M2612-34
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 61 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B12(58-62)

Quant Time: Jun 08 23:13:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 17:38:51 2021
Response via : Initial Calibration

