

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
 Data File : VU044118.D
 Acq On : 08 Jun 2021 17:20
 Operator : SY/MD
 Sample : M2580-21
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B2(0-2)

Quant Time: Jun 08 23:14:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 17:38:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	223627	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	440803	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	466114	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	223409	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	195956	50.315	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.620%	
35) Dibromofluoromethane	5.298	113	154247	49.035	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.060%	
50) Toluene-d8	7.906	98	612023	50.441	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.880%	
62) 4-Bromofluorobenzene	10.639	95	231735	48.239	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.480%	
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	42613	18.958	ug/l	99
20) Methylene Chloride	3.054	84	9458	1.146	ug/l	98
27) cis-1,2-Dichloroethene	4.678	96	5441	1.594	ug/l	85
43) Isopropyl Acetate	5.919	43	23541	2.661	ug/l	97
84) 1,2,4-Trimethylbenzene	11.475	105	3657	1.197	ug/l	91
95) Naphthalene	14.095	128	32907	4.883	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
Data File : VU044118.D
Acq On : 08 Jun 2021 17:20
Operator : SY/MD
Sample : M2580-21
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B2(0-2)

Quant Time: Jun 08 23:14:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 17:38:51 2021
Response via : Initial Calibration

