

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
 Data File : VU044120.D
 Acq On : 08 Jun 2021 18:07
 Operator : SY/MD
 Sample : M2580-24
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B3(0-2)

Quant Time: Jun 08 23:14:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 17:38:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	213142	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	420192	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	434821	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	217294	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	189563	51.067	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.140%	
35) Dibromofluoromethane	5.298	113	146059	48.710	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.420%	
50) Toluene-d8	7.906	98	568513	49.153	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.300%	
62) 4-Bromofluorobenzene	10.639	95	219097	47.846	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.700%	
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	109631	51.172	ug/l	99
20) Methylene Chloride	3.051	84	9433	1.281	ug/l	94
95) Naphthalene	14.095	128	13028	3.675	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
Data File : VU044120.D
Acq On : 08 Jun 2021 18:07
Operator : SY/MD
Sample : M2580-24
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B3(0-2)

Quant Time: Jun 08 23:14:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 17:38:51 2021
Response via : Initial Calibration

