

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
 Data File : VU044122.D
 Acq On : 08 Jun 2021 18:53
 Operator : SY/MD
 Sample : M2580-26
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B3(4-6)

Quant Time: Jun 08 23:15:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 17:38:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	220230	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	437585	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.424	117	462749	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	221107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	195988	51.099	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.200%	
35) Dibromofluoromethane	5.298	113	152855	48.950	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.900%	
50) Toluene-d8	7.906	98	601106	49.906	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.820%	
62) 4-Bromofluorobenzene	10.639	95	226720	47.542	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.080%	
Target Compounds						
					Qvalue	
16) Acetone	2.630	43	61478	27.772	ug/l	96
20) Methylene Chloride	3.054	84	10522	1.522	ug/l	93
43) Isopropyl Acetate	5.922	43	15086	1.718	ug/l	98
84) 1,2,4-Trimethylbenzene	11.475	105	4959	1.289	ug/l	91
95) Naphthalene	14.095	128	59467	6.565	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
Data File : VU044122.D
Acq On : 08 Jun 2021 18:53
Operator : SY/MD
Sample : M2580-26
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B3(4-6)

Quant Time: Jun 08 23:15:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 17:38:51 2021
Response via : Initial Calibration

