

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
 Data File : VU044126.D
 Acq On : 08 Jun 2021 20:27
 Operator : SY/MD
 Sample : M2612-37
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B12(70-74)

Quant Time: Jun 08 23:15:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 17:38:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	211063	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	424386	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	447585	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	208942	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	193054	52.520	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.040%	
35) Dibromofluoromethane	5.298	113	148817	49.139	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.280%	
50) Toluene-d8	7.906	98	591232	50.612	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.220%	
62) 4-Bromofluorobenzene	10.639	95	216604	46.834	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.660%	
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	16350	7.707	ug/l	100
20) Methylene Chloride	3.051	84	11928	2.122	ug/l	97
43) Isopropyl Acetate	5.919	43	24521	2.879	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
Data File : VU044126.D
Acq On : 08 Jun 2021 20:27
Operator : SY/MD
Sample : M2612-37
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B12(70-74)

Quant Time: Jun 08 23:15:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 17:38:51 2021
Response via : Initial Calibration

