

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
 Data File : VU044127.D
 Acq On : 08 Jun 2021 20:51
 Operator : SY/MD
 Sample : M2612-38
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 18-B12(74-78)

Quant Time: Jun 08 23:15:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 08 17:38:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	217185	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	436691	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	464589	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	216611	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	198785	52.555	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.100%	
35) Dibromofluoromethane	5.298	113	154709	49.645	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.280%	
50) Toluene-d8	7.906	98	610041	50.751	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.500%	
62) 4-Bromofluorobenzene	10.639	95	224898	47.257	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.520%	
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	16426	7.524	ug/l	94
20) Methylene Chloride	3.054	84	11794	1.970	ug/l	94
43) Isopropyl Acetate	5.919	43	25462	2.905	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060821\
Data File : VU044127.D
Acq On : 08 Jun 2021 20:51
Operator : SY/MD
Sample : M2612-38
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
18-B12(74-78)

Quant Time: Jun 08 23:15:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U060721W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 08 17:38:51 2021
Response via : Initial Calibration

